


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *MP-MV-15-CN-PN-05-M-GATEW*

S.N.: *3673*

Date: *MAR 3 1 87*



Ⓐ 40 db ATTEN  
1.0 MHz



Ⓑ 40 db ATTEN  
1.0 MHz

- a) Output signal amplitude:  
*0 TO ±5 VOLTS.*
- b) Pulse width:  
*0.2 TO 4.0 NS*
- c) Rise time:  
*≤ 60 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240*  
*50-60 Hz*