


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AN-1-C-N-OT-M*

S.N.: *3675*

Date: *MAR 20 87*



40 dB ATTEN
10 VOLTS/DIV
1.0 MHz

- a) Output signal amplitude:
0 TO - 10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz