


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-3-PS-P-PN*

S.N.: *3628*

Date: *APRIL 23 87*



40 dB ATTEN

10 VOLTS/DIV

500 KHz

- a) Output signal amplitude:
0 TO ± 20 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 200 PSEZ
- d) Fall time:
≤ 300 PSEZ
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz