


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANM-2-C-D-M-OT-P*

S.N.: *3690*

Date: *APRIL 8 '87*



- a) Output signal amplitude:  
*0 TO +15V*
- b) Pulse width:  
*0.2 TO 2.0 nSec*
- c) Rise time:  
*≤ 100 pSec*
- d) Fall time:  
*≤ 130 pSec*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*

*40 db atten*

*10. 0.0253/DIV*