


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PM-1-P*
S.N.: *3691*
Date: *APRIL 28 87*



40 dB ATTEN
5 VOLT/DIV
1.0 nSEC

- a) Output signal amplitude:
0 TO +5 VOLTS.
- b) Pulse width:
0.2 TO 6.0 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK

g) Prime power:
+24 ± 1.0 VOLT.
500 mA (MAX)
[Signature]