


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVAMP-3.-C-PN

S.N.: 3698

Date: MAY 6 87



40 dB ATTEN  
10 VOLTS/DIV  
1.0 MHz

- a) Output signal amplitude:  
0 TO ±20 VOLTS
- b) Pulse width:  
5 TO 100 NS
- c) Rise time:  
≤ 200 PSEC
- d) Fall time:  
≤ 300 PSEC
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

120 / 240 V  
50-60 Hz