


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PRR-1-PS-CA-P*

S.N.: *3705*

Date: *MAY 5 67*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
0.1 TO 5.0 μS
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

R_c = 50 Ω
1 KHz

120/240 V
50-60 Hz
AA