


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AMP-PM-HV3-C-P-PN-OT-M*

S.N.: *3706*

Date: *MAY 15 '87*



Ⓐ 50 dB ATTEN  
∴ ≈ 32 VOLTS/DIV  
1.0 MHz



Ⓑ 50 dB ATTEN  
∴ ≈ 32 VOLTS/DIV  
1.0 MHz

- a) Output signal amplitude:  
*0 TO ± 40 VOLTS.*
- b) Pulse width:  
*0.4 TO 2.0 NS*
- c) Rise time:  
*≤ 150 PSEZ*
- d) Fall time:  
*≤ 250 PSEZ*
- e) PRF:  
*0 TO 1.0 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 MHz.*  
*[Signature]*