


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AN-2-P-05-PJ*

S.N.: *3708*

Date: *FEB 10 87*



- d) Output signal amplitude:
0 TO +10 VOLTS
- e) Pulse width:
5 TO 50 NS
- f) Rise time:
≤ 100 PSEC
- g) Fall time:
≤ 135 PSEC
- h) PRF:
0 TO 50 KHz
- i) Jitter, stability:
0.1%
- j) Prime power:
120/240 J
50-60 Hz

40 db ATTEN
10 VOLTS/DIV
50 KHz