


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-7-C-PN-CA-EW-RD1


S.N.: 3715

Date: APRIL 6 '87



P out.
 $R_L = 1 \Omega$
5 KHz

- a) Output signal amplitude:
0 TO ± 10 AMPS TO
- b) Pulse width: 1-2 μ .
0.4 TO 200 μ SEC.
- c) Rise time:
 ≤ 200 PSEC
- d) Fall time:
 ≤ 200 NSEC.
- e) PRF:
MIN PW: 0 TO 5 KHz
MAX PW: 0 TO 50 Hz
- f) Jitter, stability:
OK



P out.
 $R_L = 1 \Omega$
50 Hz


- g) Prime power:
120 / 240 V
50 - 60 Hz

PULSE GENERATOR
PERFORMANCE CHECK


Model:

S.N.: 3715 CONT.

Date:



N_{out} $R_L = 2 \Omega$
5 kHz



N_{out} $R_L = 2 \Omega$
50 Hz

- a) Output signal amplitude:
- b) Pulse width:
- c) Rise time:
- d) Fall time:
- e) PRF:
- f) Jitter, stability:
- g) Prime power: