


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-2.C-NOT-M1*

S.N.: *3718*

Date: *MAY 22 7 67*



*40 db ATTEN*  
*10 VOLTS/DIV*  
*1.0 MHz*

- a) Output signal amplitude:  
*0 TO - 10 VOLTS*
- b) Pulse width:  
*5 TO 100 NS\**
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50-60 Hz*

*\* FOR PRECISE PW CONTROL NEAR 0 NS PW POT MAY BE CHANGED FROM 5K TO 500 Ω.*