


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVM-3-C-P-OT-M*

S.N.: *3721*

Date: *MAY 26 87*



① *40 db ATTEN*
10 VOLTS/DIV
≈ 12 MHz



② *AS 1A BUT*
1 NS/DIV

- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
2 TO 15 NS
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 MHz