


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMH-PS-P*

S.N.: *3729*

Date: *JUNE 4 87*



50 dB ATTEN
= 32 VOLTS/DIV

- a) Output signal amplitude:
0 TO +30 VOLTS
 - b) Pulse width:
≤ 1.0 NSEC
 - c) Rise time:
NA
 - d) Fall time:
NA
 - e) PRF:
0 TO 1.0 MHz
 - f) Jitter, stability:
OK
 - g) Prime power:
120/240V
50-60Hz.
- [Signature]*