


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-PS-N

S.N.: 3730

Date: APRIL 7 87



70 db ATTEN  
320 VOLTS/DIV  
2 KHz.

- a) Output signal amplitude:  
0 TO 350 V.
- b) Pulse width:  
5 TO 100 NS OR
- c) Rise time:  
≤ 1.5 NS OR
- d) Fall time:  
≤ 3 NS OR
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 Hz

*[Signature]*