


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-B3-P-W*
S.N.: *3732*
Date: *JUNE 3 87*



R_i = 50 Ω
5 KHz

- a) Output signal amplitude:
0 TO +100 VOLT
- b) Pulse width:
0.1 TO 5.0 μ SEC.
- c) Rise time:
 ≤ 10 NSEC
- d) Fall time:
 ≤ 10 NSEC
- e) PRF:
0 TO 5 KHz.
- f) Jitter, stability:
OK
- g) Prime power:
+24 \pm 1.0 VOLT
400 mA (MAX)

A