


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMP-MU-15-C-OT-M-P-PN*

S.N.: *3735*

Date: *JUNE 1 87*



Ⓐ 40 dB ATTEN
5 VOLTS/DIV
100 MHz



Ⓑ AS Ⓐ BUT
100 PSEC/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 4.0 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 133 PSEC
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
30-80 Hz

[Signature]