


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AUG-2-C-P  
S.N.: 3736  
Date: JUNE 4 87



60 db ATTEN.  
100 VOLTS/DIV  
50 KHz.

- a) Output signal amplitude:  
0 TO +120 VOLTS
- b) Pulse width:  
≤ 2 NSEL.
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 100 KHz.
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240V  
50-60 Hz

*[Handwritten signature]*