


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-AM-PW-C-P*

S.N.: *3738*

Date: *MAY 25 87*



- a) Output signal amplitude:  
*0 TO +200 VOLTS*
- b) Pulse width:  
*50 NSR TO 1.0 USR*
- c) Rise time:  
*≤ 10 NSR*
- d) Fall time:  
*≤ 10 NSR*
- e) PRF:  
*MIN PW: 0 TO 100 KHz*  
*MAX PW: 0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 KHz*  
*R<sub>L</sub> = 50 Ω*

*120/240 V*  
*50-60 Hz*

*[Signature]*