


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-A-2-PS-P*

S.N.: *3740*

Date: *JUNE 17 87*



*$R_L = 50 \Omega$
500 kHz*

- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width:
20 TO 200 NS
- c) Rise time:
 ≤ 5 NS
- d) Fall time:
 ≤ 5 NS
- e) PRF:
MIN PW: 0 TO 10 MHz
MAX PW: 0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz