


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-N-PN-D-OT-A*

S.N.: *3748*

Date: *JULY 21 87*



40 db ATTEN

5 VOLTS/DIV

*≈ 12 MHz.
1.8V*

- a) Output signal amplitude:
0 TO 5V
- b) Pulse width:
0.2 TO 0.6 nS
- c) Rise time:
≤ 100 pS
- d) Fall time:
≤ 135 pS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240 V
50-60 Hz*