


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-V-HV1-C-M-PN*

S.N.: *3764*

Date: *JUNE 15 87*



*ⓐ 50 db ATTEN
= 32 VOLTS/DIV
70 KHz*



*ⓑ AS ⓐ BUT
1 nSEC/DIV
(RISE TIME)*

- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
2 TO 100 nSEC
- c) Rise time:
≤ 250 pSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*

[Signature]