


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APR-1PW-C-PN-05*
 S.N.: *3766*
 Date: *JUNE 22 87*



P_{out}
R_L = 50Ω
1 kHz

- a) Output signal amplitude:
0 TO ±200 VOLTS
- b) Pulse width:
0.1 TO 5.0 USEC
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
0 TO 1.0 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz

[Signature]



P_{out} *R_L = 50Ω*
1 kHz