


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMA-3-C-N*

S.N.: *3771*

Date: *JUNE 9 87*



*40 db ATTEN*

*10 VOLTS/DIV*

*10 MHz*

- a) Output signal amplitude:  
*0 to 15 V*
- b) Pulse width:  
*2 to 15 nsec*
- c) Rise time:  
*≤ 150 psec*
- d) Fall time:  
*≤ 2 nsec*
- e) PRF:  
*0 to 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:


*120/240 V*

*50-60 Hz*

*[Signature]*

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-3-C-P*  
S.N.: *3771 (mod)*  
Date: *NOV 5 1987*



*40 db ATTEN*  
*10 VOLTS/DIV*  
*~ 10 MHz*

- a) Output signal amplitude:  
*0 TO +15 V.*
- b) Pulse width:  
*2.0 TO 15 NS*
- c) Rise time:  
*≤ 150 PSEC*
- d) Fall time:  
*≤ 600 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120 / 240 V*  
*50 60 Hz*