



PULSE GENERATOR
PERFORMANCE CHECK

Model: 192R-3-PW-PS-PN
 S.N.: 3774
 Date: JULY 7 87



P_{out} $R_L = 50 \Omega$
 1.0 KHz



N_{out} $R_L = 50 \Omega$
 50 Hz

- a) Output signal amplitude:
0 TO $\pm 200 \mu\text{V}$
- b) Pulse width:
0.1 TO $100 \mu\text{SEC}$
- c) Rise time:
 $\leq 10 \text{ nSEC}$
- d) Fall time:
 $\leq 10 \text{ nSEC}$
- e) PRF:
MIN PW: 0 TO 1.0 KHz
MAX PW: 0 TO 50 Hz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 Hz

[Handwritten signature]