


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANP-AU-MB-P
S.N.: 3775
Date: JUNE 8 87



50 dB ATTEN
32 VOLTS/DIV
100 KHZ.

- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0.4 TO 2.0 NSEC
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 250 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS
300 mA (MAX)