


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-TN-1*
S.N.: *3776*
Date: *JUNE 12 87*



≈ 60 db ATTEN
100 VOLT/DIV

- a) Output signal amplitude:
-230 V
(+30 V DC OFFSET)
- b) Pulse width:
5 TO 100 NSER.
- c) Rise time:
62 NSER
- d) Fall time:
63 NSER
- e) PRF:
0 TO 5 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240V
50-60 Hz.