


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-1-N*
S.N.: *3786*
Date: *JUN 7 8 87*



R_L = 50 Ω
100 kHz

- a) Output signal amplitude:
0 TO - 200 V.
- b) Pulse width:
0.1 TO 5.0 μ SEC.
- c) Rise time:
 \leq 10 NSEC
- d) Fall time:
 \leq 10 NSEC.
- e) PRF:
0 TO 1.0 kHz.
- f) Jitter, stability:
OK
- g) Prime power:
+24 \pm 1.0 VDC
500 mA (MAX)

[Signature]