


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVL-2A-C-PN-05-AM-W*

S.N.: *3787*

Date: *MAY 28 57*



60 db ATTEN
100 VOLTS/DIV
5 KHz

- a) Output signal amplitude:
0 TO ±170 VOLTS
- b) Pulse width:
*5 TO 400 NSEC**
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz

** NEEDS ADDITIONAL
EXTERNAL DELAY
LINE FOR 100 TO
400 NSEC RANGE.
(SEE INSTRUCTIONS)*