


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVMAN-2-C-PN*

S.N.: *3790*

Date: *FEB 30 '87*



*20 db ATTEN
① ∴ 2 VOLTS/DIV
= 12 MHz*



*③ 20 db ATTEN
≈ 50 MHz*

- a) Output signal amplitude:
0 TO ± 2 VOLTS
- b) Pulse width:
0.3 TO 5.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 150 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*

[Signature]