


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APD-2A-N*

S.N.: *3794*

Date: *Jul 14 '81*



*~7 AMPS/DIV*  
*20 KHz*

- a) Output signal amplitude:  
*0 TO -5 AMPS*
  - b) Pulse width:  
*2 TO 50 NS*
  - c) Rise time:  
*≤ 1 NS*
  - d) Fall time:  
*≤ 2 NS*
  - e) PRF:  
*0 TO 20 KHz*
  - f) Jitter, stability:  
*OK*
  - g) Prime power:  
*+15 ± 1.0 V*  
*350 mA*  
*(MAX)*
- [Signature]*