


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-V-HVI-P*

S.N.: *3796*

Date: *JUN 20 87*



Ⓐ 50 db ATTEN
20 kHz



Ⓑ 50 db ATTEN
20 kHz
(RISE TIME)

- a) Output signal amplitude:
0 TO +50 V
- b) Pulse width:
2 TO 100 NSEC
- c) Rise time:
≤ 250 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 V
250 mA (MAX)

AB