


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVL-2A-C-CMA*
S.N.: *3797 (UPGRADED FROM AVI-V-C)*
Date: *MAY 11 '89*



- a) Output signal amplitude: *0 TO +170 V.*
- b) Pulse width: *(+150 V AT 5 TO 100 NS @ 20 KHz)*
- c) Rise time: *≤ 1.5 NS @*
- d) Fall time: *≤ 3 NS @*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*

60 db ATTEN
100 VOLT/DIV
PRF = 20 KHz

50-60 Hz