


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *APR-A-1-PW-F*

S.N.: *B799*

Date: *JUN 14 '67*



*PRF = 5 KHz*  
*R<sub>L</sub> = 50Ω*



*PRF = 50 KHz*  
*R<sub>L</sub> = 50Ω*

- a) Output signal amplitude:  
*0 TO +200 V.*
- b) Pulse width:  
*50 nSEC TO 1.0 μSEC*
- c) Rise time:  
*≤ 10 nSEC*
- d) Fall time:  
*≤ 10 nSEC*
- e) PRF:  
*MIN PW 0 TO 100 KHz*  
*MAX PW 0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 ± 1.0 VOLTS*  
*600 mA (MAX)*

*[Signature]*