



PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANMP-3-C-CA-P*
S.N.: *3802*
Date: *JULY 22 '87*



*(A) 40 db ATTEN
10 VOLTS/DIV
1.0 MHz*



*(B) AS (A) BUT
RISE TIME*

- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 300 PSEZ
- d) Fall time:
≤ 300 PSEZ
- e) PRF:
0 TO 100 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz.*

[Signature]