


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PA-AN-AVI-C-N-05*

S.N.: *3803*

Date: *JULY 28 87*



40 dB ATTEN
1.0 VOLTS/DIV
1.0 MHz

- a) Output signal amplitude:
0 TO 15 V
- b) Pulse width:
0.3 TO 2 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 150 PSEC
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz

[Signature]