


PULSE GENERATOR
PERFORMANCE CHECK


Model: AMMN-1-C-P-PN-MA-0T

S.N.: 3821

Date: JULY 23 87



Ⓐ 40 dB ATTEN
-.5 VOLTS/DIV
≈ 12 MHz



Ⓑ AS Ⓐ BUT
≈ 50 MHz

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.3 TO 5.0 NSER
- c) Rise time:
≤ 100 PSEER
- d) Fall time:
≤ 200 PSEER
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 MHz