


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-TN-1

S.N.: 3822

Date: JUNE 11 87



= 60 db ATTEN  
= 100 VOLTS/DIV  
2 KHz

- a) Output signal amplitude:  
-230 VOLTS
- b) Pulse width:  
5 TO 100 NSEC
- c) Rise time:  
≤ 2 NSEC
- d) Fall time:  
≤ 3 NSEC
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

120/240 V  
50-60 Hz

*[Signature]*