


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVMP-2 PS PW*

S.N.: *3832*

Date: *JUN 22 87*



(1) *40 db ATTEN*
10 VOLTS/DIV
1.0 MHZ



(2) *AS (1) BUT*
RISE TIME.

a) Output signal amplitude:

0 TO ± 10 VOLTS

b) Pulse width:

5 TO 100 NS

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 1.0 MHZ

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz

[Signature]