


PULSE GENERATOR
PERFORMANCE CHECK


Model: *ARK-A-0-PW*

S.N.: *3834*


Date: *JULY 31 '87*



(A) *40 db ATTEN
MIN PW*



(B) *40 db ATTEN
RISE TIME*



(C) *40 db ATTEN*

- a) Output signal amplitude:
0 TO ±15 VOLTS
- b) Pulse width:
4 TO 100 NSER
- c) Rise time:
≤ 100 NSER
- d) Fall time:
≤ 2 NSER
- e) PRF:
MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
170/240 V
50-60 Hz

[Signature]