



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVN-1-C-PN-M*
 S.N.: *3835*
 Date: *JUNE 11 87*



(A) 40 db ATTEN
 5 VOLTS/DIV
 ≈ 250 MHz



(B) AS @ BUT
 ≈ 50 MHz

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time:
≤ 100 PSEK
- d) Fall time:
≤ 100 PSEK
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
- g) Prime power:

OK
120/240 V
50-60 Hz.
[Signature]