



PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-AV-1-C-A  
S.N.: 3838  
Date: JUN 7 31 87



Ⓐ 60 db ATTEN  
100 VOLT/DIV  
5.0 KHZ



Ⓑ AS Ⓐ BUT  
1 VOLT/DIV  
(RISE TIME)

- a) Output signal amplitude:  
0 TO 400 V.
- b) Pulse width:  
5 TO 100 NSER
- c) Rise time:  
≤ 1 NSER
- d) Fall time:  
≤ 3 NSER
- e) PRF:  
0 TO 5 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
120 / 240 V  
50-60 HZ  
JK