


PULSE GENERATOR
PERFORMANCE CHECK


Model: ACAMP-2-C-M-EP1

S.N.: 3840

Date: JULY 17 87



Ⓐ 40 db ATTEN
10 VOLTS/DIV
3.0 MHz



Ⓑ 1V Ⓐ BUT
1 MHz/DIV

- a) Output signal amplitude:
0 TO +10 V.
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 3.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz

[Signature]