


PULSE GENERATOR
PERFORMANCE CHECK

Model: *BNH MPT-P*
S.N.: *3844*
Date: *Aug 5 80*



- a) Output signal amplitude:
≥ 30 VOLTS.
- b) Pulse width:
≤ 1 nSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLTS
150mA (MAX)

[Signature]

50 db ATTEN

32 VOLTS/DIV

500 KHz.