



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AM4-AM2-P*  
 S.N.: *3850*  
 Date: *AMAY 26 87*



Ⓐ 50 db ATTEN  
 ≈ 32 VOLTS/DIV  
 500 MHz



Ⓑ 100 ps/DIV BUT  
 100 pSEC/DIV  
 (JITTER CHECK)

- a) Output signal amplitude:  
*> + 60 VOLTS*
- b) Pulse width:  
*≤ 2 nSEC.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:
- g) Prime power:  
*OK*  
*+135 ± 5 VOLTS*  
*5 mA (MAX)*

*[Signature]*