



PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-B1-P*
S.N.: *3855*
Date: *APR 26 '87*



Ⓐ $R_L = 50\Omega$
 100 kHz



Ⓑ $R_L = 50\Omega$
 100 kHz

- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
20 TO 200 NSEC
- c) Rise time:
 ≤ 3 NSEC
- d) Fall time:
 ≤ 3 NSEC
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 \pm 1.0 VOLT
300 mA (MAX)

[Signature]