


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APV-AV-1-CPN*

S.N.: *3860*

Date: *Jul 6 87*



- a) Output signal amplitude:
0 TO ±10 VDC
- b) Pulse width:
0.2 TO 4.0 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz

40 db ATTEN
10 VOLTS/DIV
1.0 MHz.