


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVB1-300-C*  
S.N.: *3876*  
Date: *AUG 18 87*



*50 dB ATTEN*  
*32 VOLTS/DIV*  
*100 KHz*

- a) Output signal amplitude: *≥ 100 V<sub>pp</sub>*
- b) ~~Pulse width:~~  
*PERIOD: 300 nS*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*  
*50-60 Hz*