

PULSE GENERATOR

PERFORMANCE CHECK

Model:


ANL-AN-1-C-T-W-AN-05

S.N.:


3881

Date:


SEPT 14 87



① -60 dB ATTEN
100 VOLTS/DIV
5 MHz



② AS ① BUT 1.0 NS/DIV
(RISE TIME)



③ AS A BUT NOUT,
MAX TR.

a) Output signal amplitude:

0 TO ±100 VOLTS

b) Pulse width:

5 TO 400 NS^{*}

c) Rise time:

≤ 1 NS TO 10 NS

d) Fall time:

≤ 3 NS

e) PRF:

0 TO 5 MHz

f) Jitter, stability:

OK

g) Prime power:

120/240 VOLTS

50-60 Hz