


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVT-V-HV3-C-N*

S.N.: *3882*

Date: *Aug 26 87*



*(A) 60 db ATTEN  
100 V/DIV/DIV  
5 KHz*



*(B) AS (A) BUT  
1 NS/DIV  
(RISE TIME)*

- a) Output signal amplitude:  
*0 TO - 100 V.*
- b) Pulse width:  
*4 TO 100 NS*
- c) Rise time:  
*≤ 200 PSEC*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/280 V  
50-60 Hz.*

*[Signature]*