


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-1-PS-7-05*
S.N.: *3891*
Date: *AUG 31 87*



R_L = 50 Ω
1.0 MHz

- a) Output signal amplitude:
0 TO +200 V.
- b) Pulse width:
0.1 TO 5.0 μ SEC
- c) Rise time:
 \leq 10 nSEC
- d) Fall time:
 \leq 10 nSEC
- e) PRF:
0 TO 1.0 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.
- h) OFFSET
0 TO \pm 50 VOLTS.
250 mA (MAX)